

DECLARATION OF CONFORMITY

Application of Council EMC Directive 89/336/EEC.

NOTE

The Low Voltage Directive 73/23/EEC does not apply to the Model 2701-B PaveTracker Plus gauge because there are no voltages greater than 75 V dc that are accessible by the operator.

Standards to which Conformity is Declared:

EN 55011:1998+A1:1999+A2:2002 – Industrial, Scientific, and Medical (ISM) Radio Frequency Equipment


EN 61326:1997+A1:1998+A2:2001+A3:2002 – Electrical Equipment for Measurement, Control, and Laboratory Use

Manufacturer: Troxler Electronic Laboratories, Inc.
P.O. Box 12057, 3008 Cornwallis Road
Research Triangle Park, North Carolina
27709 U.S.A.

Apparatus: Model 2701-B PaveTracker Plus

Year of Declaration: 2006

Signature:



Date:

5 May 2006

Full Name:

William F. Troxler, Jr.

Title:

President & Chairman of the Board

Place of issue:

Research Triangle Park, NC, U.S.A.